


PULSE GENERATOR
PERFORMANCE CHECK

Model: AP-M-1A2-C-P-PT



S.N.: 9717

Date: MAR 29 2001



- a) Output signal amplitude:
0 TO +14 VOLTS
- b) Pulse width:
0 TO 2 NS
- c) Rise time:
≤ 300 ps
- d) Fall time:
≤ 300 ps
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power: 120/240V
50-60Hz

Ⓜ 40db ATTN ∴ 10V/DIV



RISE TIME (A)