



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-2-e-P*
 S.N.: *9725*
 Date: *MAR 21 2001*



- a) Output signal amplitude:
0 TO +10 AMPS
- b) Pulse width:
2 TO 15 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60Hz



[Handwritten signature]