


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-5B*

S.N.: *9738*

Date: *APRIL 20 2001*



a) Output signal amplitude:

0 to +30V

b) Pulse width:

c) Rise time: *≤ 350ps (20-80%)*
NA

d) Fall time:

NA

e) PRF: *0 to 1 MHz*

f) Jitter, stability:

OK

g) Prime power:

+15VDC

200mA MAX

40dB ATTEN
- 10V/DIVS-
PRF ≈ 300KHz