


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5B
S.N.: 9739
Date: APRIL 20 2001



- a) Output signal amplitude:
0 to +30 V
- b) Pulse width:
≤ 350 ps (20-80%)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTN
10V/DIV
PRF ≈ 300 kHz

+ 15 VDC
200 mA (MAX)