


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANM-3-C-N-M1

S.N.: 9756

Date: Aug 1 2001



- a) Output signal amplitude:
0 to -15V (to 50ns)
- b) Pulse width:
0 to 30ns (PRF < 7.5 MHz)
- c) Rise time:
0 to 20ns (7.5 ≤ PRF ≤ 10 MHz)
≤ 150 ps
- d) Fall time:
≤ 600 ps.
- e) PRF:
0 to 10 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

40 dB return

10 V/div / DIV

TOP 20 ns / DIV

BOT 500 ps / DIV

PRF = 10 MHz

100 ± 10V

50-60 MHz