


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-1-C-PN-AT*

S.N.: *9757*

Date: *JUNE 14 2001*



- a) Output signal amplitude:
0 TO ± 200 V (TO 50N)
- b) Pulse width:
① *2 TO 5 NS*
- c) Rise time:
≤ 0.8 NS
- ② d) Fall time:
≤ 1 NS
- e) PRF:
③ *0 TO 50 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTN

100 VOLTS/DIV

PRF ≈ 20 KHz

- ① *PW MAX, 2 NS/DIV*
- ② *PW MIN, 2 NS/DIV*
- ③ *1 NS/DIV*

100 ± 10 VOLTS
50 - 60 MHz