


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANP-AU-1A3-P
S.N.: 9776
Date: JUL 25 2001



- a) Output signal amplitude:
0 to +40 V (to 50Ω)
- b) Pulse width:
0.4 to 2.0 ns
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250 ps
- e) PRF: 0 to 1 MHz
- f) Jitter, stability: OK
- g) Prime power: +15 VDC.
300 mA (max)

50 dB ATTN
≈ 32 VOLTS / DIV
PRF = 100 KHz