


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PMMH-4-C-P*

S.N.: *9798*

Date: *JUNE 11 2001*



- a) Output signal amplitude:  
*0.70 + 100 VOLTS*
- b) Pulse width:  
*0.50 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF: *0.70 10 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*

*60 dB ATTEN*  
*∴ 100 VOLTS/DIV*  
*TOP PRF = 1 MHz*  
*BOT PRF = 10 MHz*

*120/240V*  
*50-60 Hz*  
*A*