


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV14-5B

S.N.: 9813

Date: APRIL 20 2001



- a) Output signal amplitude:
0 to +30V.
- b) Pulse width:
≤ 350 ps (20-80%)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: 0 to 1.0 MHz.
- f) Jitter, stability: OK
- g) Prime power:

40 dB ATTN

10 VOLTS/DIV

PRF = 300 kHz

+ 15 VDC

200 mA (MAX)

~~NA~~