


PULSE GENERATOR
PERFORMANCE CHECK

Model: AW-9A-C-F
 S.N.: 9818
 Date: JUNE 26 2001



- a) Output signal amplitude:
0 TO +200 mA
- b) Pulse width:
0.4 TO 4.0 NS
- c) Rise time:
≤ 200ps
- d) Fall time:
≤ 200ps
- e) PRF:
0 TO 1.0 MHz

TOP MAIN FLAME OUT
 TO 40db. NO AX-51
 10V/DIV, 1NS/DIV
 PRF = 100 KHz

- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50 60Hz

BOT AS TOP BUT AX-51
 INSTALLED + ME OUT
 TO 20db. IN 459A DOSE
 USED AS A SIMULATED LOAD

JA