


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVN-1-C-N-PN-M-OT

S.N.: 9854

Date: Aug 16 2001



- a) Output signal amplitude:
0.70 ± 5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 135ps
- e) PRF: 50 - 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 lb ATTN
5 VOLTS/DIV

TOP 1 NS/DIV

BOI 100 PS/DIV
(RISE TIME)

120/248V
50-60 Hz