


PULSE GENERATOR
PERFORMANCE CHECK

Model: *NI-V-C-P*
S.N.: *9055*
Date: *AUG 20 2001*



- a) Output signal amplitude:
0 TO +50V (7050V)
- b) Pulse width:
2.0 TO 100NS
- c) Rise time:
≤ 350 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20KHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V

50-60 Hz
A

50 dB ATTN
32 VOLTS/DIV.
TOP 10NS/DIV
PROI. 500 pA/DIV
(RISE TIME)
PRF = 20KHz.