


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-AV-15-C-OT-8*

S.N.: *9869*

Date: *Jul 15 2001*



- a) Output signal amplitude:
0 to +5 VOLT (70 50Ω)
- b) Pulse width:
0.2 to 4.0 NS
- c) Rise time:
≤ 60 ps
- d) Fall time:
≤ 135 ps
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50 100W

40 db ATTEN
-.5 VOLT/DIV
PRF = 100 kHz

TOP 1 NS/DIV

BOT 100 ps/DIV (RISE TIME)
(20 80%)

UY