


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-SB-C-P

S.N.: 9914

Date: SEPT 6 2001



- a) Output signal amplitude:
0 to +30V (70.50mV)
- b) Pulse width:
≤ 350ps
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: 0 to 1.0MHz
- f) Jitter, stability:
OK
- g) Prime power:
170/240V

40dB ATTN
∴ 10 VOLTS/DIV

TOP 1.0 NS/DIV

BOT 200 ps/DIV

PRF ≈ 100 KHz.

50-60dB