


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14-1401-P*
S.N.: *9949*
Date: *NOV 2 2001*



- a) Output signal amplitude:
0 TO +100 V (70.50.50)
- b) Pulse width:
≤ 1 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 100 KHz.*
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN
100 VOLT/DIV.
PRF ≈ 50 KHz.

+15 VDC
250 mA (MAX)
[Signature]