


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1-P*
 S.N.: *9954*
 Date: *OCT 15 2001*



40 db ATTN.: 10 VOLTS/DIV.

TOP 1.0 NS/DIV
BOT 100 ps/DIV
(RISE TIME)

PRF = 100 kHz

- a) Output signal amplitude:
0 TO +10V (70.50%)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC.
250 mA (MAX)

[Signature]