


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-MU-1-P*

S.N.: *9955*

Date: *OCT 15 2001*



40dB ATTEN. : 10 V/DIV

TDP 1.0 NS/DIV

*BUT 100ps/DIV
(RISE TIME)*

PRF = 100 kHz

- a) Output signal amplitude:
0 TO +10 V (TO 50V)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps.
- d) Fall time:
≤ 135 ps.
- e) PRF:
0 TO 1.0 MHz.
- f) Jitter, stability:
OK.
- g) Prime power:
+15 VDC
250 mA (MAX)