


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MP-AU-1-P*

S.N.: *9956*

Date: *04 15 2001*



- a) Output signal amplitude:  
*0 TO +10 V (TO 50%)*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 135 ps*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 dB ATTEN.: 10V/DIV*

*TOP 1.0 NS/DIV*

*BOT 100 ps/DIV  
(RISE TIME)*

*100 KHz PRF*

*+15 VDC  
250 mA (MAX)*

*[Signature]*