


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-B-1P*  
S.N.: *9963*  
Date: *SEPT 11 2001*



- a) Output signal amplitude:  
*0 TO +350 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF: *0 TO 5 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*

*60 dB ATTEN*  
*100 VOLTS/DIV*  
*TOP 10 NS/DIV*  
*BOT 2 NS/DIV (RISETIME)*  
*PRF = 2.0 KHz.*

*50-60 Hz*  
*[Signature]*