


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-2-C-N-OT-EO-M*

S.N.: *9966*

Date: *APR 22 2001*



a) Output signal amplitude:

0 TO -15 VOLTS

b) Pulse width:

0.2 TO 2.0 NS

c) Rise time:

≤ 100 ps.

d) Fall time:

≤ 135 ps.

e) PRF:

0 TO 25 MHz.

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz.

b) DC OFFSET

0 TO ± 5 VOLTS.

i) MONITOR ✓ (20 dB)

40 dB ATTN
10 VOLTS/DIV

TOP 1.0 NS/DIV

BOT 100 ps/DIV
(RISE TIME)

PRF ≈ 12 MHz.