


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-EGGIE*

S.N.: *9980*

Date: *OCT 16 - 2007*



- a) Output signal amplitude: ✓
≥ +10 V_{DC} (TO5:3)
- b) Pulse width: ✓
1.5 NS FWHM
- c) Rise time: ✓
≤ 100 ps -
- d) Fall time: ✓
≈ 3 NS -
- e) PRF: ✓
0 TO 2 MHz -
- f) Jitter, stability: ✓
OK -
- g) Prime power:

a) *+15 VDC -*
160 mA

b) *-5 VDC*
10 mA

[Signature]

40 db ATTEN
10 V_{DC} / DIV

TOP: 1 NS / DIV

BOT: 100 ps / DIV

PRF = 2.0 MHz