


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MH-4V1-P-095-M1*

S.N.: *9996*

Date: *NOV 23 2001*



- a) Output signal amplitude: *0 to +100 V (CTB 50%)*
- b) Pulse width: *≤ 1 ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 200 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 V DC*

*60 dB ATTN*

*100 V/DIV*

*PRF = 200 kHz*

*250 mA (max)*