

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV 706A-PS-1-M-KMPIA*

S.N.: *9999*

Date: *NOV 28 2001*

- a) Output signal amplitude:
0 TO +40 AMPMS TO
- b) Pulse width:
0 TO +100µS
- c) Rise time:
1.0 TO 100µS
(10% MAX DUTY CYCLE)
≤ 0.5µS
- d) Fall time:
≤ 0.5µS
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
(10% MAX DUTY CYCLE)
OK
- g) Prime power:
120/240V
50-60 Hz.



9999

$$R_L = .205$$

$I_{max} (\approx 40 \text{ AMPS})$

LOAD
VOLTAGE
MONITOR
VOLTAGE

