


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-M-1-P-M-EN*

S.N.: *10015*

Date: *04 26 2001*



- a) Output signal amplitude:
0 to +10V (to 5.0V)
- b) Pulse width:
0.2 to 4.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 to 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
250 mA MAX.

40 db ATTN

10 V/DIV

TAP 1 ns/DIV

BOT 100 ps/DIV
(RISE TIME)

PRF = 100 kHz