


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-3-EA-M-05-P*

S.N.: *10026*

Date: *DEC 20 2001*



- a) Output signal amplitude:  
*0 to +500V (70.50s)*
- b) Pulse width:  
*≤ 5 NS (HALF POWER)*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 10 NS (20-80%)*
- e) PRF:  
*0 TO 20 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VDC  
400 mA (MAX)*
- h) EA : *OK ✓*
- i) M : *OK ✓*
- j) OS : *OK ✓*

*60 db ATTN*

*100 VOLT/DIV*

*PRF ≈ 10 KHz.*