


PULSE GENERATOR
PERFORMANCE CHECK

Model: *198-MV-1-N*
S.N.: *10045*
Date: *DEC 6 2001*



- a) Output signal amplitude:
0 TO 10V (TO 50N)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps.
- d) Fall time:
≤ 135 ps.
- e) PRF:
0 TO 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
250 mA (MAX)

40 db ATTN
10 VOLTS/DIV
TOP 1.0 NS/DIV
BOT 100 ps/DIV
(RISE TIME)
100 KHz PRF