

PULSE GENERATOR

PERFORMANCE CHECK

Model: AN-106A-PS-P-M-R5-KMP1A

S.N.: 10050

Date: DEC 12 2001

- a) Output signal amplitude:
0 TO +40 AMPS TO
0 TO +10 VOLTS
- b) Pulse width:
1.0 TO 100US
- c) Rise time:
(1% max DUTY CYCLE)
≤ 0.5 US
- d) Fall time:
≤ 0.5 US
- e) PRF: 0 TO 1 KHz.
(1% max DUTY CYCLE)
- f) Jitter, stability:
6K
- g) Prime power:
120/240V
50-60 HZ.



10050

$R_c \approx 0.205$

$I_{max} (\approx 40 \text{ Amp})$

LOAD
VOLTAGE

INSTRUMENT
VOLTAGE

