


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMR-2-C-PN*
S.N.: *10051*
Date: *FEB 22 2002*



- a) Output signal amplitude: *0 TO ± 200 V (TO 50V)*
- b) Pulse width: *2 TO 10 NS*
- c) Rise time: *6 NS*
- d) Fall time: *6.2 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *170/2000 V*

60 dB ATTEN
100 Volts/DIV
PRF = 20 KHz

80-60 V
[Signature]