


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-MV-1V3-CP*

S.N.: *10061*

Date: *DEC 18 2001*



- a) Output signal amplitude:
0 TO +40 V (TO 50V)
- b) Pulse width:
0.4 TO 2.0NS
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/2500 V
30-60 MHz

[Signature]

50 db ATTN
-32 V/DIV
TOP 1NS/DIV
BOT 200 ps/DIV

PRF = 100 kHz