


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANN-1-C-P*

S.N.: *10078*

Date: *DEC 18 2001*



- a) Output signal amplitude: *0.70 ± 5 VOUT*
- b) Pulse width: *0.270 10 NS*
- c) Rise time: *6 100 ps*
- d) Fall time: *6 135 ps*
- e) PRF: *50 to 250 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50 - 60 Hz

40 dV ATTEN

5 V/DIV

TOP 1 ns/DIV

MOT 200 ps/DIV