


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PNH-5-1-C-P-KMP1*

S.N.: *10032*

Date: *MAR 6 2002*



- a) Output signal amplitude: *0.70 + 10 VOLT (10.50V)*
- b) Pulse width: *≤ 130 ps.*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0.70 / 1 MHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240 VOLT*
50 - 60 MHz.

40 dB ATTN
- 10 VOLT/DIV
TOP 1 NS/DIV
BOT. 200 ps/DIV
PRF = 100 KHz.