


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-N*

S.N.: *10089*

Date: *DEC 14 2001*



- a) Output signal amplitude:
0 TO -5 V (TO 50%)
- b) Pulse width:
0.2 TO 6 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 pW @ 50-60 MHz

40 db ATTN

-5 VOLTS/DIV

PRF ≈ 12 MHz

TOP 1 ns/DIV

BOT 100 ps/DIV (RISE TIME)