


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PM-1-C-N-PN*

S.N.: *10090*

Date: *DEC 14 2001*



- a) Output signal amplitude:  
*0 TO ±5 VOLTS (P 50N)*
- b) Pulse width:  
*0.2 TO 6.0 NS*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 135 ps*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 MHz*

*40 dB ATTN*  
*5 VOLTS/DIV*  
*TOP 1 NS/DIV*  
*BOT 100 ps/DIV (RISE TIME)*  
*PRF ≈ 12 MHz*