


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-3-P-2A*

S.N.: *10105*

Date: *MAR 13 2002*



a) Output signal amplitude:

*0 TO +15 VOLTS*

b) Pulse width:

*0 TO 15NS*

c) Rise time:

*≤ 150ps*

d) Fall time:

*≤ 600ps*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+30 TO +36 VDC  
(500 mA MAX)*

*40 dB ATTN*

*10 VOLTS/DIV*

*TOP 2 NS/DIV*

*BOT 200ps/DIV  
(RISE TIME)*