


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-5-1-c-f*

S.N.: *10111*

Date: *FEB 28 2002*



- a) Output signal amplitude: *0 to +10V (70.50n)*
- b) Pulse width: *≤ 130ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

40 dB ATTEN
10 VOLT/ DIV
PRF = 1.0 MHz

120/240V
50 to 0Hz
[Signature]