


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-1-PS-D*

S.N.: *10112*

Date: *FEB 21 2002*



- a) Output signal amplitude:  
*0 TO +10V (TO 50V)*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 135 ps*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz.*

*[Signature]*