


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-PS-P*

S.N.: *10113*

Date: *FEB 21 2002*



- a) Output signal amplitude:
0 TO +10V (50Ω)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240Vrms
50-60 MHz
[Signature]