


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVPP-1-C-P-HUTA


S.N.: 10118

Date: MAR 5 2002



- a) Output signal amplitude: 0 TO +10 VOLTS (TO 50 mA)
- b) Pulse width: 0.2 TO 100 NS
- c) Rise time: ≤ 100 ps
- d) Fall time: ≤ 135 ps
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240 V
50-60 Hz
- h) A-B DELTA
0 TO 25 NS.

Ⓐ TOP 500 ps/DIV
BOT 5.0 NS/DIV



TOP: 500 ps/DIV
BOT: 5 NS/DIV
PRF = 100 KHz
40 dB ATTEN.