


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-143CP-PS-EKA

S.N.: 10119


Date: February 8, 2002

“Out A” Test



Top: IN, +4V pulse, 2V / div, 200 ns / div
Bottom: OUT, +30V pulse into 50Ω, 10V / div

“Out B” Test



Top: IN, +4V pulse, 2V / div, 200 ns / div
Bottom: OUT, +30V pulse into 50Ω, 10V / div

a) Output Signal Amplitude: 0 to +30V

b) Gain: +7.5 V/V, nominal

c) Rise Time: < 50 ns

d) Fall Time: < 50 ns

e) PRF: N/A

f) Jitter, Stability: OK

g) Prime Power: 120/240V, 50-60 Hz