


PULSE GENERATOR

PERFORMANCE CHECK

Model: PVL-2-C-N-0A

S.N.: 1021

Date: MAR 21 2002



- a) Output signal amplitude:  
0 to 350 V (to 50n)
- b) Pulse width:  
5 to 100 ns.
- c) Rise time:  
 $\leq 2$  ns
- d) Fall time:  
 $\leq 3$  ns
- e) PRF:  
0 to 5 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

120/200 V  
50-60 kHz.

*[Signature]*

60 dB ATTEN

100 V/10ns/DIV

PRF = 2.0 kHz