


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-52-C-07A*

S.N.: *10125*

Date: *MAR 10 2002*



- a) Output signal amplitude:
0 to ±75 VOLTS TO
- b) Pulse width: *(50 DIV)*
10 NS (FWHM)
- c) Rise time:
≤ 2.5 NS
- d) Fall time:
≤ 2.5 NS
- e) PRF: *0 TO 3.09 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/200V
50-60Hz