


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-3A-P-6A*

S.N.: *10136*

Date: *FEB 12 2002*



- a) Output signal amplitude:
0 TO +450V
- b) Pulse width:
≤ 4NS (20% point)
- c) Rise time:
≤ 1NS
- d) Fall time:
≤ 2.5NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN

-100 VOLTS/DIV

PRF = 10 KHz.

+15 VDC

70 mA (max)