


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APM-1-PS-PPN-ET-ON-ED*

S.N.: *10177*

Date: *APRIL 29 2002*



- a) Output signal amplitude: *0 TO 15 VOLTS*
- b) Pulse width: *0 TO 50 ns*
0.2 TO 6 ns
- c) Rise time: ≤ 100 ps
- d) Fall time: ≤ 135 ps
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/200 V*
50-60 MHz

40 dB ATTN
10 VOLTS DIV
TOP 1 NS/DIV
150T 100 ps/DIV (RISE TIME)
PRF = 10 MHz

120/200 V
50-60 MHz
[Signature]