


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3-C-P-PN-D-M-OT-ECL*

S.N.: *10184*

Date: *APRIL 22 2002*



- a) Output signal amplitude:
0 TO ±15 VOLTS
- b) Pulse width: *UPDL = 50n*
2 TO 15 NS
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 600 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 VOLTS

50-60 MHz

[Signature]

40 dB ATTEN

10 VOLTS/DIV

TOP 5 NS/DIV

BOT. 200 ps/DIV (RISE TIME)

PRF = 12 MHz