


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-5-1-C-P*

S.N.: *10185*

Date: *MAY 7 2002*



- a) Output signal amplitude: *0 to +10 Volts (to 50s)*
- b) Pulse width: *≤ 130 ps.*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 1 MHz.*
- f) Jitter, stability: *OK.*
- g) Prime power: *120 pW*
50 to 100 dB.

[Signature]