


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MPP-1-C-P-DN-BA-EN*

S.N.: *10183*

Date: *MAR 19 2002*



a) Output signal amplitude:  
*0 TO ± 10 VOLTS (TO 50V)*

b) Pulse width:  
*0.2 TO 100 NS*

*1 NS/DIV* Rise time: *≤ 100 NS*

*1 NS/DIV* d) Fall time: *≤ 135 PS*

e) PRF: *0 TO 1.0 MHz*

*5 NS/DIV* f) Jitter, stability: *OK*

g) Prime power: *100 ± 10 VOLTS*  
*50 - 60 MHz*

*40 dB ATTEN*  
*10 VOLTS / DIV*  
*NAF = 100K Hz*