


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVUMR-1-N*
S.N.: *10191*
Date: *MAY 29 2002*



- a) Output signal amplitude:
0 TO -5 VOLTS (TO 50V)
- b) Pulse width:
10 NS TO 200 NS
- c) Rise time *(20% MAX DUTY CYCLE)*
≤ 1 NS
- d) Fall time:
≤ 1 NS
- e) PRF: *0 TO 10 MHz*
- f) Jitter, stability: *(20% MAX DUTY CYCLE)*
OK.
- g) Prime power:
+24 VDC
200 mA (MAX)

40 dB ATTN.
-5 VOLTS / DIV.
PRF = 1 MHz.