


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-1-T*
S.N.: *10192*
Date: *MAY 02 02*



- a) Output signal amplitude:
0 TO +5 V (7050V)
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 ns
- d) Fall time:
≤ 135 ns
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC

500 mA (max)

40 dB ATTEN
-.5 VOLTS/DIV
TOP 1 NS/DIV
BUT 100ps/DIV (RISE TIME)
PRF = 10 MHz