


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM14-4-C-P-M*

S.N.: *10206*

Date: *MAY 17 2002*



- a) Output signal amplitude: *0 TO +100 V. (7050 Ω)*
- b) Pulse width: *$\leq 1NS$*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*

60 dB ATTEN.

100 VOLTS/DIV

PRF = 5 MHz.

50-60 Hz