



P.O. BOX 265
OGDENSBURG, NY
U.S.A. 13669-0265

TEL: 888-670-8729 (USA & Canada) or +1-613-226-5772 (Intl)
FAX: 800-561-1970 (USA & Canada) or +1-613-226-2802 (Intl)


BOX 5120, LCD MERIVALE
OTTAWA, ONTARIO
CANADA K2C 3H4

info@avtechpulse.com - http://www.avtechpulse.com/

PERFORMANCE CHECKSHEET

Model: AVR-8A-B-P
Type: High Voltage Pulse Generator
S.N.: 10212 (repair)
Date: July 28, 2006

Rise & Fall Time Test



10 Hz, 2 us, +1 kV into a 50Ω load.
200 V/div, 400 ns/div.

a) Output Signal Amplitude: 0 to +1000V

b) Pulse Width (FWHM): 200 ns - 200 us,
0.2% duty cycle

c) Rise Time (20%-80%): < 100 ns


d) Fall Time (20%-80%): < 100 ns

e) PRF: 1 Hz - 1 kHz

f) Jitter, Stability: OK

g) Prime Power: 100-240V AC, 50-60 Hz.

Wide Pulse Test



10 Hz, 100 us, +1 kV into a 50Ω load.
200 V/div, 20 us/div.

28 Jul 2006
10:18:29